



Certificate No. 0790/025666

**"EC DESIGN-EXAMINATION"**  
**CERTIFICATE IN ACCORDANCE WITH MODULE B1 OF THE**  
**PRESSURE EQUIPMENT DIRECTIVE**

**SGS Reference No.**  
**PC/45667/B1/01**

**Manufacturer:**

Name : Sechrist Industries Inc.  
Address : 4225 E. La Palma Avenue, Anaheim, California 92807

SGS United Kingdom  
Limited  
Johns Lane,  
Tividale,  
Oldbury,  
West Midlands  
B69 3HX

**Pressure Equipment Description: Model 3300 Monoplace Hyperbaric Oxygen Chamber**

Drawing numbers : 190004 Rev. 2  
Volume : 1200 litres  
Design pressure : 2.43 bar  
Design temperature : 38 °C  
Contents : 100% Oxygen  
Corrosion allowance : Nil  
Test pressure : 3.66 bar

The undersigned certifies that the design of this pressure equipment conforms with the requirements of the Pressure Equipment Directive (97/23/EC).

This certificate is valid for 10 years and is renewable on request. Valid until : 20/02/2018

Date : 20<sup>th</sup> February 2008

Name : M . J . V . Bartlett

Position : Verification Engineer

Notified body identification number : 0353

Signature :

